



Certificate of Calibration

Asset Serial Number: 303000
Calibration Type: Pressure
Certification Date: July 17, 2019
Calibration Range: 0 to 1003 dBar
Sensor Range: 0 to 1000 dBar
Residual (RMSE): 0.040 dBar
Standards: Paro 785

Coefficients

Coefficient A:	-1.253833E+2	Coefficient H:	0.000000E+0
Coefficient B:	0.000000E+0	Coefficient I:	-3.831245E-10
Coefficient C:	0.000000E+0	Coefficient J:	0.000000E+0
Coefficient D:	0.000000E+0	Coefficient K:	0.000000E+0
Coefficient E:	1.918260E-2	Coefficient L:	0.000000E+0
Coefficient F:	0.000000E+0	Coefficient M:	-1.221923E-14
Coefficient G:	0.000000E+0	Coefficient N:	0.000000E+0

Robert Haydock
President, AML Oceanographic

AML Oceanographic certifies that the asset described above has been calibrated or recalibrated with equipment referenced to traceable standards. If this instrument or sensor has been re-calibrated, please be sure to update your records. Please also ensure that you update the instrument's coefficient values in any post-processing software that you use, if necessary.