



Certificate of Calibration

Asset Serial Number: 400503
Calibration Type: Temperature
Certification Date: January 03, 2019
Calibration Range: -1.9 to 32 °C
Sensor Range: -2 to +32 °C
Residual (RMSE): 0.001 °C
Standards: Hart 1560/3471

Coefficients

Coefficient A:	-8.622933E+0	Coefficient H:	0.000000E+0
Coefficient B:	1.492043E-3	Coefficient I:	0.000000E+0
Coefficient C:	-3.216191E-8	Coefficient J:	0.000000E+0
Coefficient D:	1.047605E-12	Coefficient K:	0.000000E+0
Coefficient E:	-2.278428E-17	Coefficient L:	0.000000E+0
Coefficient F:	3.021455E-22	Coefficient M:	0.000000E+0
Coefficient G:	-1.662227E-27	Coefficient N:	0.000000E+0

Robert Haydock
President, AML Oceanographic

AML Oceanographic certifies that the asset described above has been calibrated or recalibrated with equipment referenced to traceable standards. If this instrument or sensor has been re-calibrated, please be sure to update your records. Please also ensure that you update the instrument's coefficient values in any post-processing software that you use, if necessary.